

August 16, 2023

VI 2345121:1-4 **Coliform Bacteria Analysis**

Customer ID : 4020266

Water System Solutions

South Fork Estates MWC

Attn: Thomas Ridenour

403 Scranton Ave.

Porterville, CA 93257

System Number : CA5403113

Project Name : Monthly Bacteriological

Analytical Results

ID	Sample Description	Total	Fecal	E. Coli	Units	Method	Prep	Footnote
1	Well 02	Absent	---	Absent	A/P/100ml	SM 9223B	Colilert-P/A 18	T
2	Storage Tank	Absent	---	Absent	A/P/100ml	SM 9223B	Colilert-P/A 18	
3	Quail Run	Absent	---	Absent	A/P/100ml	SM 9223B	Colilert-P/A 18	
4	Black Oaks	Absent	---	Absent	A/P/100ml	SM 9223B	Colilert-P/A 18	

N/R Not Required

MPN Most Probable Number

A/P Absence/Presence

T Hold time violation

The samples listed above were Acceptable for both Total and Fecal Coliform

Sample Handling Information

ID	Sample Number	System Number	Sample Type/Reason	Sampler	Employed By	Sampled
1	VI 2345121-001	CA5403113_002_002	Source-Other	Thomas Ridenour	Water Systems Soluti	2023-08-05 07:00
2	VI 2345121-002	5403113	System-Other	Thomas Ridenour	Water Systems Soluti	2023-08-05 07:15
3	VI 2345121-003	5403113	System-Other	Thomas Ridenour	Water Systems Soluti	2023-08-05 07:10
4	VI 2345121-004	5403113	System-Routine	Thomas Ridenour	Water Systems Soluti	2023-08-05 07:20

Field Analysis/QA Information

ID	Sample Description	Cl Total/Free mg/l	Temp	Analysis Started	Analysis Completed	Contact	Contacted
1	Well 02	---/---	---	2023-08-05 15:17 ios	2023-08-06 10:31 ios	N/R	
2	Storage Tank	---/---	---	2023-08-05 15:17 ios	2023-08-06 10:31 ios	N/R	
3	Quail Run	---/---	---	2023-08-05 15:17 ios	2023-08-06 10:31 ios	N/R	
4	Black Oaks	---/---	---	2023-08-05 15:17 ios	2023-08-06 10:31 ios	N/R	

Analyses were performed at the FGL Visalia Laboratory using Standard Methods 20th edition. If you have any questions regarding your results, please call. The FGL Visalia Laboratory is certified by California ELAP #2810 and accredited to ISO/IEC 17025:2017 by PJLA certificate #75605, Testing.

Prepared By: RRH

cc:SDHS-CDPH FRESNO-B

Reviewed and
Approved By

Raquel R. Harvey



Digitally signed by Raquel R. Harvey
Title: Tech Director Microbiology
Date: 2023-08-16

42527:08/07/2023 TEST DESCRIPTION - See Reverse side for Container, Preservative and Sampling information

Client: WSS- South Fork Estates MWC
Address: Water System Solutions
South Fork Estates MWC
Attn: Thomas Ridenour
403 Scranton Ave.
Porterville, CA 93257

Phone: (559)359-7172 Fax:
Contact Person: Thomas Ridenour

Project Name: **Monthly Bacteriological**

Purchase Order Number:
Quote Number:

Sampler(s) *Thomas Ridenour*

Sampling Fee: _____ Pickup Fee: _____

Compositor Setup Date: ___/___/___ Time: ___/___

Lab Number: VI *2345121* 4-20266

Samp Num	Location Description	Date Sampled	Time Sampled	Method of Sampling:	Type of Sample	Potable(P)	Non-Potable(NP)	Ag Water(AgW)	Bacteri Type: Other(O)	System(SYS)	Source(SR)	Waste(W)	Bacteri Reason: Routine(ROUT)	Repeat(RPT)	Replace(RPL)	Other(O)	Special(SPL)	Coliform-Colilert-P/A	120ml(PBa)-Na2S2O3
1	Well 02	8/5/23	7:00Am	G	DW	P			SR	SPL								1	
2	Storage Tank		7:15Am	G	DW				Sys	SPL								1	
3	Quail Run		7:10Am	G	DW				1	ROUT								1	
4	Black Oaks		7:20Am	G	DW					SPL								1	

Remarks:	Relinquished <i>[Signature]</i> Date: 8/5/23 Time: 1334	Relinquished Date: Time:	Relinquished Date: Time:
	Received By: <i>[Signature]</i> Date: 8-5-2023 Time: 1334	Received By: Date: Time:	Received By: Date: Time:

Corporate Offices & Laboratory
853 Corporation Street
Santa Paula, CA 93060
Phone: (805) 392-2000
Env Fax: (805) 525-4172 / Ag Fax: (805) 392-2063

Office & Laboratory
2500 Stagecoach Road
Stockton, CA 95215
Phone: (209) 942-0182
Fax: (209) 942-0423

Office & Laboratory
563 E. Lindo
Chico, CA 95926
Phone: (530) 343-5818
Fax: (530) 343-3807

Office & Laboratory
3442 Empresa Drive, Suite D
San Luis Obispo, CA 93401
Phone: (805) 783-2940
Fax: (805) 783-2912

Office & Laboratory
9415 W. Goshen Avenue
Visalia, CA 93291
Phone: (559) 734-9473
Fax: (559) 734-8435

TH 407

9.8°C
ROI

/